

Total Acid Number of marine engine oils

Summary Information

- Correlates well with ASTM D664
- Detection range: 0.01–99 mg KOH/g
- Portable and battery operated
- Mid-Infrared Spectroscopy

Product Description

The IR Sphinx spectrometer measure the mid-infrared spectrum of a sample and extract relevant parameters. The spectrometers do not contain any moving parts but use a solid state dispersion element in combination with black body infrared emitters to measure the infrared spectrum of a sample. This results in a unique product which is robust, battery operated and weighs less than 0.5 kg. The spectrometer can be configured to measure from 2.5 μm –5.0 μm or from 5.5 μm –11.0 μm . The IR Sphinx spectrometer come with a sophisticated but user friendly software called Sphinx Suite. The software is modular and the user can choose from a number of different software modules. The software is compatible with many common operating systems.

Application

The combustion process in engines produces acidic by-products which are absorbed by the engine oils. The Total Acid Number (TAN) is a measure of the acidity of the oil. The acidity or the TAN of the oil has to be controlled as acidic components corrode the metallic parts of the engine. In modern engine oils the acidic components are neutralised by alkaline components which are described by the Total Base Number (TBN). By monitoring the TAN of the oil important information about the quality and performance of the oil can be extracted.

The ASTM standard to determine the Total Acid Number is described in ASTM D664 and it is based on a titration method. However the Total Acid Number can also be accurately measured using mid-infrared spectroscopy and the results correlates well with the relevant ASTM standard.

How to use

The IR Sphinx spectrometer enable the user to quickly measure the Total Base Number of an oil sample. Depending on the product range the measurement is carried out in a slightly different way. For the IR Sphinx ATR products the oil sample placed on top of the ATR crystal making sure that the entire crystal is covered by the oil.


For the IR Sphinx transmission products the oil sample has to be present in the sample chamber. The transmission systems are best suited for inline measurement where the oil sample is delivered to the sample holder via a pumping system. Alternatively a syringe can be used to deliver the sample to the sample holder.

Once the sample is in place the measurement is started from the software. After about 30s the analysis of the sample is available.


Comline Elektronik Elektrotechnik GmbH
Karl-Rapp-Straße 1 • 92442 Wackersdorf • Germany
Fon +49 94 31 / 75 65-0 • Fax 75 65-25
comline@comline-elektronik.de • www.comline-elektronik.de



Results & Performance



Mid-infrared absorption spectra of a typical marine engine oil at different points during the lifetime of the oil. Note the visible changes in the spectra which are related to specific oil degradation effects.



The Total Acid Number (TAN) has been calculated from the measured absorption spectra using the SphinxSuite software module. The plot compares the TAN obtained from analysing the mid-infrared absorption spectra to the TAN obtained from the same oil sample in an accredited oil analysis laboratory. The correlation between the results is exceptional.

Comline
Sensor Systems